

MAYOR
Randy A. Garrison

CITY CLERK/CONTROLLER
Rita S. Lee



CITY COUNCIL
Virginia Alexander
Ken Doss
Charles "Chuck" Gill
Dwight Tankersley
Kenny Thompson

November 2, 2020

REQUEST FOR BID

The City of Hartselle is seeking bids for **(3) Cardiac Monitor/Defibrillators**

Individual sealed bids must be submitted by **Thursday, November 18, 2020 at 3:15 p.m. CST**, at which time they will be opened and publicly read aloud. Additional copies of the bid specifications and bid form can be obtained from the City Clerk's Office between the hours of 7:30 a.m. and 4:30 p.m. Monday through Friday. **Bids should be submitted to:**

City of Hartselle

ATTN: Rita S. Lee, City Clerk

200 Sparkman Street N.W.

Hartselle, AL 35640

All bid envelopes, including courier envelopes inside and outside, should be clearly marked as containing a "Sealed bid for Fire Department Cardiac Monitors."

Attached is a copy of the minimum specifications required by the City of Hartselle. Should you have any questions you may contact **Daryl Fox, Fire Chief, Hartselle Fire & Rescue at (256) 751-4945.**

The City of Hartselle has the right to accept or reject any or all bids and to waive any technicalities.

No bidder may withdraw his bid for a period of thirty (30) days after the time set for receiving bids.

OFFICE 256-773-2535
200 Sparkman Street NW

FAX 256-773-2257
Hartselle, Alabama 35640

www.hartselle.org

All cardiac monitors/ defibrillators and all components shall be biomedical recertified and/or new. All monitors shall be in production by the manufacture. Each shall include the following:

- 12-Lead ECG
- Biphasic
- Pacing
- AED
- SpO2/SpCO
- NiBP
- EtCO2
- Bluetooth
- Defibrillator
- 2 new batteries
- Carrying case and/or accessory storage
- AC power adapter that includes right angle plug to connect unit to a 110v outlet

The cardiac monitors/ defibrillators shall come with all necessary accessories for immediate operation out of the box with up-to-date software. All of the equipment shall be biomedical recertified and/or new units that come with a 1-year warranty provided and to included accessories.

Each unit that is not brand new shall be biomedical recertified, be cleaned and tested by biomedical technicians who follows the manufacturer’s checklist to confirm that the unit is operational and working as intended.

The AC Power Adapter allows for continuously charge your device batteries without the need to swap them out ensuring this unit is always operational.

This Bid is for (3) cardiac monitors/defibrillators with all options listed above.

Company Name: _____

Representative: Print _____

Sign _____

Bid \$ _____